



169 Lexington Court, Unit i
Waterloo, ON Canada, N2J 4R9
Toll Free: (877) 746-4764
Local: (519) 746-4761
Email: info@indigoinstruments.com

Certificate of Analysis

Product Code: [33813-Hard](#)

Product Description: Hard Water Test Strips

Each lot of test strips is tested both for precision and accuracy

The Water Hardness test strip is calibrated at 0, 50, 125, 250, and 500ppm and is also calibrated at 0, 3, 7, 15, and 30 grains per gallon. The Water Hardness test strip measures water hardness due to calcium carbonate.

Precision Testing

10 strips from each lot are tested using standards prepared from ACS grade chemicals.

Strips are evaluated on a pass/fail basis.

All 10 strips must pass (all develop color with little or no variation in response) for the lot to be acceptable.

Accuracy Testing

Standards are prepared by serial dilution of stock solutions.

Standards are prepared at values of 50, 125, 250, and 500ppm

Each standard is prepared using ACS grade chemicals and deionized water.

Following the label instruction, test strips are tested using each standard with results determined by comparing the color developed by the strip against an approved color chart.

Test strips are accepted if the color developed for the standard being evaluated visually falls within $\pm 1/2$ color unit.

The test strips are guaranteed to perform to specification for a minimum of two years from the date of manufacture (lot) date provided the test strips are stored in the original container and kept in a cool, dry environment out of direct sunlight.