



Stability Statement

Part Number: [33814-VitC](#)

ProductDescription: Vitamin C Test Strips

The test strips should be stored in the closed bag in a cool, dry place out of direct sunlight. The ideal condition is 25C (77F). A range of 20C (68F) - 30C (86F) is acceptable. Test strips should not be stored in a refrigerator or freezer. Test strips should not be stored where they can be exposed to heat (for example, inside a vehicle or shed), damp or high humidity environments.

When stored properly, the test papers have a minimum shelf-life of 3 years from the date of manufacture. Under proper storage conditions the strips will continue to perform properly beyond the expiration date.

Indigo Instruments

169 Lexington Crt. Unit i

Waterloo, ON N2J 4R9

Phone: (519) 746 4761

Email: info@indigoinstruments.com

Web: www.indigoinstruments.com