



# Sialkot Material Testing Laboratory

UNIT ESTABLISHED BY TRADE DEVELOPMENT AUTHORITY OF PAKISTAN MINISTRY OF COMMERCE GOVT. OF PAKISTAN  
NEAR SUBLIME CHOWK, MARALA ROAD, SIALKOT - 51310 - PAKISTAN  
TEL: 0092-52-3562818 FAX: 0092-52-3554217



Serial No. **1062**.....

Ref. No. PM/SIMTEL/12/868  
December 6, 2012

Receipt No. ....1502....

## ANALYTICAL / TEST REPORT

In Respect of : Hyssop Surgical Sialkot  
Sample Received from : Representative  
Name of the Sample : Scissors  
Type/Grade of the Material : AISI-420  
Specification against which tested : Analytical Results Got on Direct Emission Spectrometer

### Results are in Percentage:

|                 |                |
|-----------------|----------------|
| Cr ..... 12.485 |                |
| C ..... 0.197   | Mn ..... 0.521 |
| P ..... 0.015   | S ..... 0.021  |
| Si ..... 0.269  | Ni ..... 0.101 |

This analytical /test report is based solely on the particular sample(s) supplied by the party and should not be reproduced in part.

Sampling has not been performed by the SIMTEL and the SIMTEL does not accept responsibility that samples(s) supplied is/are truly representative samples(s) of any batch or stock or entire production.

While the SIMTEL agrees to take every reasonable precaution to ensure validity of its test results, it assumes no liability for the accuracy thereof beyond the amount of the fee charged for the analysis/test.

The party shall assume full responsibility for the ethical use of the results in the analysis/test reports and the laboratory shall be held free from any and all claims, which may result from the use of such data by the party or others.

After completion of the report the sample will be preserved for three month until negotiated otherwise.



Engr. Amjad Hussain  
Project Director

*Sialkot  
Material  
Testing  
Laboratory*